



CSITC Task Force Contributions

Axel Drieling
Bremen Fibre Institute (FIBRE) /
ICA Bremen



22nd Meeting of the CSITC Task Force Thessaloniki, Greece, November 2, 2014



A. Drieling: CSITC Contributions

Contributions Thessaloniki, 2014-11







Development of CSITC RT Participation and Lab Evaluation Results

2


A. Drieling: CSITC Contributions


Thessaloniki, 2014-11






	Number of Participants Median Evaluations								
•		Participating Labs	Evaluation Combined Prop.	Evaluation Micronaire •	Evaluation Strength •	Evaluation Length •	Evaluation Uniformity •	Evaluation Color Rd 🕶	Evaluation Color +b
Scale Value				0.1	1.5	0.02	1	1.5	0.5
AV 2007-14	108.4	71.5	0.49	0.49	0.44	0.40	0.37	0.47	0.47
AV 2007-11	94.8	65.5	0.51	0.51	0.47	0.41	0.37	0.50	0.49
AV 2013-14	134.3	85.7	0.44	0.45	0.38	0.38	0.38	0.39	0.41
2013-1	114	76	0.47	0.40	0.39	0.35	0.40	0.41	0.49
2013-2	129	84	0.44	0.43	0.38	0.38	0.36	0.47	0.47
2013-3	148	90	0.45	0.40	0.36	0.37	0.38	0.47	0.53
2013-4	151	93	0.45	0.44	0.41	0.39	0.37	0.39	0.44
2014-1	123	82	0.40	0.48	0.34	0.39	0.31	0.28	0.27
2014-2	128	86	0.45	0.52	0.35	0.43	0.42	0.35	0.33
2014-3	147	89	0.44	0.51	0.45	0.33	0.39	0.36	0.36
2014-4									









Within Limit Evaluation based on 1 test per sample



	% of instruments >=95% within limits									
		based on	single tests	per sample						
RT	Mic Str Len: UHML L-Unif Col: Rd									
Limits	0.2	2	0.03	2	1.5	1				
Average	85.9	57.7	70.1	81.8	60.4	85.8				
2011-3	90	57	74	84	61	82				
2011-4	85	53	61	76	63	81				
2012-1	87	52	70	74	61	85				
2012-2	90	63	76	79	59	85				
2012-3	88	61	78	90	60	80				
2013-1	87	54	72	78	50	85				
2013-2	86	57	69	81	52	87				
2013-3	85	58	70	82	50	84				
2013-4	85	50	70	85	63	90				
2014-1	84	63	71	88	67	95				
2014-2	83	69	61	82	71	88				
2014-3	82	56	68	82	69	87				
2014-4										

No improvements visible

Commercially important results that include varation betweeen single tests.

9

A. Drieling: CSITC Contributions

Thessaloniki, 2014-11







Comparison of RT Results to USDA Established Results


10


A. Drieling: CSITC Contributions

Thessaloniki, 2014-11











Comparison to USDA Established CFC/CAC/33 project was co-funded by the European Union and the Common Find for Common Find for



- Conclusions
 - Following CSITC RT averages (which are taken for evaluating laboratories) for the same bale shows that results are very stable
 - If there are general systematic deviations from the USDA Established results (valid for all bales), this can be detected
 - · Example: Micronaire
 - With no general systematic deviation, still there can be deviations for each bale. The results show that the RT averages are very stable, perhaps even more than the USDA Established results.
 - The decision for choosing RT averages as reference for evaluating labs, was and is the best choice.
 - Next analysis to do: detect systematic deviations in the result range

A. Drieling: CSITC Contributions





Differences between Instrument Types

- Deviations From RT to RT a)
- b) Deviations analyzing the result range
- Inter-instrument variation
- Within-instrument variation d)

Based on RTs 2013-1 to 2014-3

Statistics are strongly influenced by different number of participating instruments

A. Drieling: CSITC Contributions

Thessaloniki, 2014-11





Micronaire	Number of instr.	AV result	AV - Diff.	SD inter-instr.	SD Within- instr. Between tests	SD Within-instr. Between days
Others	5					
Premier	·					
ART+HFT	13	4.23	-0.033	0.082	0.044	0.031
ART2	9	4.24	-0.019	0.074	0.059	0.044
Uster						
HVI 1000	64	4.28	0.019	0.047	0.034	0.029
Spectrum	18	4.24	-0.023	0.074	0.050	0.034
HVI 900	21	4.25	-0.012	0.063	0.043	0.036
All	130	4.26	0.000	0.066	0.041	0.032

Variation between instruments: <= 0.03

Bias over the range: no

Differences in within-instrument variation (betw. tests):

Differences in inter-instrument variation (betw. days):

17 A. Drieling: CSITC Contributions Thessaloniki, 2014-11





Strength	Number of instr.	AV result	AV - Diff.	SD inter-instr.	SD Within- instr. Between tests	SD Within-instr. Between days
Others	6					
Premier						
ART+HFT	13	28.62	-0.06	0.91	0.55	0.37
ART2	9	28.77	0.10	0.84	0.50	0.39
Uster						
HVI 1000	64	28.68	0.01	0.66	0.55	0.40
Spectrum	18	28.53	-0.14	0.87	0.64	0.50
HVI 900	20	28.82	0.15	0.92	0.66	0.48
All	130	28.67	0.00	0.82	0.58	0.43

Variation between instruments: <= 0.15 g/tex

Bias over the range: no

Differences in within-instrument variation (betw. tests):

Differences in inter-instrument variation (betw. days):

18 A. Drieling: CSITC Contributions Thessaloniki, 2014-11





Length	Number of instr.	AV result	AV - Diff.	SD inter-instr.	SD Within- instr. Between tests	SD Within-instr. Between days
Others	6					
Premier						
ART+HFT	13	1.085	-0.002	0.013	0.013	0.008
ART2	9	1.087	0.000	0.010	0.011	0.007
Uster						
HVI 1000	64	1.087	0.000	0.009	0.009	0.006
Spectrum	19	1.086	-0.002	0.012	0.011	0.007
HVI 900	20	1.090	0.003	0.012	0.013	0.008
All	130	1.087	0.000	0.011	0.011	0.007

Variation between instruments: <= 0.003

Bias over the range: no

Differences in within-instrument variation (betw. tests):

Differences in inter-instrument variation (betw. days):

19 A. Drieling: CSITC Contributions Thessaloniki, 2014-11



Length Uniformity





	Number			SD	SD Within- instr. Between	SD Within-instr.
L. Uniformity	of instr.	AV result	AV - Diff.	inter-instr.	tests	Between days
Others	6					
Premier						
ART+HFT	13	81.19	-0.02	0.59	0.58	0.36
ART2	9	81.29	0.07	0.56	0.51	0.36
Uster						
HVI 1000	64	81.13	-0.08	0.44	0.50	0.30
Spectrum	18	81.26	0.04	0.57	0.54	0.33
HVI 900	21	81.39	0.17	0.61	0.61	0.36
All	131	81.22	0.00	0.54	0.53	0.32

Variation between instruments: <= 0.17 !!!

Bias over the range: no

Differences in within-instrument variation (betw. tests): Differences in inter-instrument variation (betw. days):

20 A. Drieling: CSITC Contributions Thessaloniki, 2014-11





Color Rd	Number of instr.	AV result	AV - Diff.	SD inter-instr.	SD Within- instr. Between tests	SD Within-instr. Between days
Others	6					
Premier						
ART+HFT	12	76.18	-0.16	0.91	0.39	0.41
ART2	9	76.04	-0.30	0.95	0.44	0.36
Uster						
HVI 1000	64	76.68	0.35	0.56	0.16	0.20
Spectrum	17	75.84	-0.50	0.89	0.29	0.36
HVI 900	21	75.98	-0.36	0.87	0.37	0.45
All	127	76.34	0.00	0.85	0.27	0.30

Variation between instruments: <= 0.5 !!!

Bias over the range: no

Differences in within-instrument variation (betw. tests):

Differences in inter-instrument variation (betw. days):

21 A. Drieling: CSITC Contributions Thessaloniki, 2014-11





Color +b	Number of instr.	AV result	AV - Diff.	SD inter-instr.	SD Within- instr. Between tests	SD Within-instr. Between days
Others	5					
Premier						
ART+HFT	11	10.75	-0.11	0.34	0.20	0.22
ART2	9	10.85	-0.01	0.31	0.21	0.16
Uster						
HVI 1000	64	10.91	0.05	0.24	0.08	0.11
Spectrum	18	10.83	-0.03	0.29	0.13	0.14
HVI 900	21	10.83	-0.03	0.28	0.12	0.16
All	127	10.87	0.00	0.29	0.12	0.14

Variation between instruments: <= 0.11
Bias over the range: perhaps

Differences in within-instrument variation (betw. tests):

Differences in inter-instrument variation (betw. days):

22 A. Drieling: CSITC Contributions Thessaloniki, 2014-11

